

TECHNICAL SPECIFICATIONS FOR Hand And Foot Unit :

1. The device should be ISO Certified or CE certified
2. The device should have Atleast 8 UVB Lamp and 8 UVA Lamp.
3. The device should Have atleast 20 Watt Power Out Put from Each lamp.
4. The device should have Option to Use Separately for Hand and Foot
5. The device should have a Control Unit for Delivering Calibrated Power
6. The device should Be provided With Atleast 5 Patient Use Googles
7. The device should Have Option to Use in Standing Position and Sitting Position
8. The Device have Option to Use UVA And UVB Separately.