



अखिल भारतीय आयुर्विज्ञान संस्थान, गोरखपुर All India Institute of Medical Sciences, Gorakhpur

कुनराघाट गोरखपुर उत्तर प्रदेश - 273008

Kunraghat, Gorakhpur, Uttar Pradesh 273008

(स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार द्वारा स्थापित एक स्वायत्त निकाय)

(An autonomous organization under the Ministry of Health & Family Welfare, Govt. of India)

Comp- 6356

Ref No. AIIMS/GKP/EW/ 571

Date :24/4/2026

Quotation Notice

Quotations are invited from intending registered Stockist/Distributors/ Suppliers/ OEMs having GST and relevant documents for supply & installation of cardiac table (**annexure attached**) at AIIMS, gorakhpur. The quotation should be dropped in the Quotation Box located at Ground Floor of Administrative Block, Medical College Building, AIIMS, Gorakhpur, U.P. latest by 30th Apr, 2026 till 4:00 pm. Any quotation received through e-mail or after the last date will not be entertained. Details of material

Terms & Condition

Prices should be clearly mentioned in the offer.

- The sealed envelope should bear quotations number and the subject of quotation (Eng. Department).
- Supply & installation should be completed within 30 days from the date of purchase order.
- Price should be inclusive of GST.
- Contact number with address and email id should be mentioned in the quotation.
- Order will be placed to the L-1 vendor on total value basis.
- Validity of the quotation should be 30 days from the date of opening and validity.
- Validity period for rates should be clearly mentioned in the quotation.
- Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
- Sample may be asked for evaluation before/after final placement of order.

Following quotations will be summarily rejected:

01. Quotations not received in sealed envelope.
02. Quotations without quotation enquiry number.
03. Quotations received without mentioning of Subject at the top of envelope.
04. Quotations received without mentioning of Name of the firm on envelope.
05. Quotations without GSTIN Number.
06. Quotations received without mentioning of address & email-ID of the firm
07. Quotations received without mentioning date of validity

For any clarifications, kindly contact at email address: se@aiimgorakhpur.edu.in.

LT Col Manish Tiwari / लेफ्टिनेंट कर्नल मनीष तिवारी
Superintendent Engineer / अधीनस्थ अभियंता
Department of Engineering / इंजीनियरिंग विभाग
All India Institute of Medical Sciences, Gorakhpur-273008
अखिल भारतीय आयुर्विज्ञान संस्थान गोरखपुर-273008
24/4/26

No.		Match	Mismatch
1	Should have anti-scratch, good surface finish ABS Laminated top having dimension 740-780 mm Lx 440-480 mm W approximately with +/-5% tolerance	✓	
2	Table top height can be adjusted with the help of operating lever which activates the gas spring. Adjustable from 760 to 1050 mm (±5%)	✓	
3	Base frame should be made of ERW round tube with 50.8 mm dia and 1.6 mm thickness. Housing should be made of aluminum extruded inner and outer tubes. Handle for gas spring made of MS sheet metal of Section 74 mm x115mm with 3 mm thick Handle with CRCA material making strong lever and providing wider area for grip	✓	
4	Castors : High endurance anti-static, Plastic injection molded castors are provided of ø50mm powder coating should be Bacteriostatic and thermosetting epoxy polyester, formulated to fulfill the requirements for bacterial protection.	✓	
5	Safe working Load : 20Kg or more	✓	
TECHNICAL STATUS		Comments received for the Department	

LT Col Manoj T. ...
 Superintendent
 Department
 All India Institute of Medical Sciences
 Anandapur, New Delhi-273008
 अखिल भारतीय आयुर्विज्ञान संस्थान, नई दिल्ली-273008



LT Col Manisha Tewari / लेफ्टिनेंट कर्नल मनिषा तिवारी
Superintendent Engineer / अधीक्षक अभियंता
Department of Engineering / इंजीनियरिंग विभाग
All India Institute of Medical Sciences, Gorakhpur-273008
अखिल भारतीय आयुर्विज्ञान संस्थान, गोरखपुर-273008