अखिल भारतीय आयुर्विज्ञान संस्थान,गोरखपुर
All India Institute of Medical Sciences, Gorakhpur
(स्वास्थ्य एवं परिवार कल्याण मंत्रालय भारत सरकार द्वारा स्थापित एक स्वायत्त निकाय)
(An autonomous organization under the Ministry of Health & Family Welfare, Govt. of India)

Quotation No.: AIIMS/GKP/COMP-7795/2026-27/2762 **Date: 26/10/2026**

Quotation Notice

Quotations are invited from intending registered stockiest /Distributors/ Suppliers/ OEMs having GST and relevant documents for **Procurement of Variable-Range Single-Channel Micropipettes for the dept. of Pathology and Laboratory Medicine** at AIIMS, Gorakhpur. The quotation should be dropped in the Quotation Box located at Ground Floor of **Administrative Block, Medical College Building, AIIMS, Gorakhpur, U.P. latest by 02/11/2026 till 16:00 PM.** Any quotation received through e-mail or after the last date will not be entertained.

Detail of the items: -

S. No.	Description	Quantity required
1	Micropipette, 0.2–2 µL	02
2	Micropipette, 1–10 µL	02
3	Micropipette, 5–50 µL	02
4	Micropipette, 10–100 µL	02
5	Micropipette, 20–200 µL	02
6	Micropipette, 100–1000 µL	02

Terms & Condition:

1. Firm to mention Make/Brand name in their quotation along with price.
2. Prices should be clearly mentioned in the offer.
3. The sealed envelope should bear quotations number and the subject of quotation.
4. Supply should be made within 15 days from the date of purchase order.
5. Price should be inclusive of freight on road charges.
6. Quotation will be evaluated on total value basis.
7. Order will be placed to the L-1 vendor on total value basis.
8. Validity of the quotation should be 30 days from the date of opening and validity period should be clearly mentioned in the quotation.
9. Sample to be submitted as and when asked for.
10. Firm to submit documentary evidence in support of claim of GST at the time of submission of bills.
11. Following quotations will be summarily rejected:
    i. Quotations not received in sealed envelope quotations without quotation enquiry number.
    ii. Quotations received without mentioning of Subject at the top of envelope
    iii. Quotations received without mentioning of Name of the firm on envelope
    iv. Quotations without GSTIN Number, Quotations without Date.
    v. Quotations without validity period. Quotations without mentioning of GST amount.
    vi. Quotations without price quoted against each item
12. The quotation should be submitted only during Working Days & Office hours i.e. Monday to Friday (9 AM to 5 PM) & Saturday (9 AM to 2 PM)
13. For any clarifications, kindly contact at email address: procurementcell@aiimsgorkhpur.edu.in
14. Specification attached as Annexure1.

S/d
Administrative officer
AIIMS Gorakhpur

Annexure - L

Specifications		
S.N.	**Item**	**GeM specifications**
1.	Manual single-channel variable volume micropipette 0.2-2 µL	• Type of micropipette: Variable volume single channel • Material of piston: PEI(Polyetherimide) • Micropipette with spring loaded tip cone: Yes • Autoclavable: Yes • Locking mechanism: Yes • Availability of digital volume indicator • Minimum volume of variable volume pipette: 0.2 µL • Maximum volume of variable volume pipette: 2 µL • Volume in increment: ~~0.002 µL~~ 0.1 µL (POINT ONE)
2.	Manual single-channel variable volume micropipette 1-10 µL	• Type of micropipette: Variable volume single channel • Material of piston: PEI(Polyetherimide) • Micropipette with spring loaded tip cone: Yes • Autoclavable: Yes • Locking mechanism: Yes • Availability of digital volume indicator • Minimum volume of variable volume pipette: 1 µL • Maximum volume of variable volume pipette: 10 µL • Volume in increment: ~~0.02 µL~~ 0.1µL
3.	Manual single-channel variable volume micropipette 5-50µL	• Type of micropipette: Variable volume single channel • Material of piston: PEI(Polyetherimide) • Micropipette with spring loaded tip cone: Yes • Autoclavable: Yes • Locking mechanism: Yes • Availability of digital volume indicator • Minimum volume of variable volume pipette: 5 µL • Maximum volume of variable volume pipette: 50 µL • Volume in increment: ~~0.1 µL~~ 1 µL
4.	Manual single-channel variable volume micropipette 10-100µL	• Type of micropipette: Variable volume single channel • Material of piston: PEI(Polyetherimide) • Micropipette with spring loaded tip cone: Yes • Autoclavable: Yes • Locking mechanism: Yes • Availability of digital volume indicator • Minimum volume of variable volume pipette: 10 µL • Maximum volume of variable volume pipette: 100 µL • Volume in increment: 1 µL
5.	Manual single-channel variable volume micropipette 20-200µL	• Type of micropipette: Variable volume single channel • Material of piston: PEI(Polyetherimide) • Micropipette with spring loaded tip cone: Yes • Autoclavable: Yes • Locking mechanism: Yes

डॉ. मिम्मा वि.एम. / Dr. Mimma V. M.
सहायक प्रोफेसर / Assistant Professor
पैथोलॉजी एवं प्रयोगशाला औषधि विभाग
Department of Pathology & Laboratory Medicine
अखिल भारतीय आयुर्विज्ञान संस्थान, गोरखपुर-273008
All India Institute of Medical Sciences, Gorakhpur-273008

		• Availability of digital volume indicator • Minimum volume of variable volume pipette: 20 µL • Maximum volume of variable volume pipette: 200 µL • Volume in increment: ~~0.2 µL~~ 1 µL
6.	Manual single-channel variable volume micropipette 100-1000µL	• Type of micropipette: Variable volume single channel • Material of piston: PEI(Polyetherimide) • Micropipette with spring loaded tip cone: Yes • Autoclavable: Yes • Locking mechanism: Yes • Availability of digital volume indicator • Minimum volume of variable volume pipette: 100 µL • Maximum volume of variable volume pipette: 1000 µL • Volume in increment: 5µL

डॉ. मिम्ना वि.एम./Dr. Mimna V. M.
सहायक प्रोफेसर/Assistant Professor
पैथालॉजी एवं प्रयोगशाला औषधि विभाग
Department of Pathology & Laboratory Medicine
अखिल भारतीय आयुर्विज्ञान संस्थान, गोरखपुर-273008
All India Institute of Medical Sciences, Gorakhpur-273008

13/6

Pg. 2/2